




操作手册



在开机前装入电池





2 x 1,5V Mignon - R6 - AA

碱性锰电池

- 2 -

MEM

显示储存的
测量值



- 5 -


MEM

打开仪器
按MEM键
最近的测量值将显
示出来

- 6 -

装入/更换电池
按测量键
进行测量
按测量键
显示储存的测
量值

RPM



在测量过程中按键

技术数据

精度

$\pm 0,02\%$
(1 digit)

操作温度

0°...50°C
(270°...323°F)
(32°...122°F)

存储温度

-20°...70°C
(253°...343°F)
(-4°...158°F)

测量时间

0,5 sec / 1 period

关机

30 sec
dates will be stored

交货范围

